


# EC Certificate

Full Quality Assurance System  
Directive 93/42/EEC on Medical Devices, Annex II excluding (4)

Registration No.: HD 2058015-1

Manufacturer: Ningbo Jiangbei Woson Medical  
Instrument Co., Ltd.  
No. 25, Lane 300, Jinshan Road,  
Jiangbei District, Ningbo  
315032 Zhejiang  
P.R. China

Products: Steam Sterilizers



TÜVRheinland®


The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II section 4 is required.

Report No.: 244310228-200

Effective date: 2021-04-21

Expiry date: 2024-05-26

Issue date: 2021-04-21



---

Jason Pan  
TÜV Rheinland LGA Products GmbH  
Tillystraße 2 · 90431 Nürnberg · Germany